

Case number **A12/00339/22**

Product (what the product is) **Particle filter mask**



Name (on the product or the packaging) **EireMed Respirator Facemask**

Brand (on the product or the packaging) **• EireMed**

OECD Portal category (if known) **85000000 - Safety / Protection - DIY**

Is the product (also) sold online? **Yes - Other**

Professional product? **No**

Product Recall

Type/number of model **KN95**

Batch number **Unknown**

Bar code **Unknown**

Product description **White PPE facemasks, branded 'EireMed'. The masks have the CE marking, the claims KN95 and FFP2, and a reference to the European standard EN149 embossed on the side.**

Packaging description **The product is sold in transparent plastic pouches, marked with FFP2 and KN95, a reference to the European standard EN149, and with the CE marking on it.**

Is the product counterfeit? **No**

Is this notification linked to a Business Gateway case?

Type of risk **Health risk / other**

Risk description (technical defect and consequences)

The particle/filter retention of the material (measured values: as low as 78%) and the total filtration capacity of the mask (measured mean values: as low as 80%) are insufficient. Consequently, the product does not fulfil the health and safety requirements; thus, even if combined with other recommended measures, it may not properly protect.

Legal provisions (at EU level) and European standards against which the product was tested and did not comply

The product does not comply with the Personal Protective Equipment (PPE) Regulation and with the European standard EN 149.

(Overall) risk level **Serious**

Product category **Protective equipment**

Measures **Voluntary measures**

Manufacturer

**Stop of sales
09/09/2021**

Year

2022

Country of origin

People's Republic of China

Alert submitted by

Ireland