

Case number **A12/00256/22**

Product (what the product is) **Particle filter mask**



Name (on the product or the packaging) **PROTECTIVE RESPIRATOR FFP2**

Brand (on the product or the packaging) **• BALTIC MASKS**

OECD Portal category (if known) **85000000 - Safety / Protection - DIY**

Is the product (also) sold online? **Unknown - Unknown**

Professional product? **No**

Product Recall

Type/number of model **BM-001V**

Batch number **BM00001V**

Bar code **4779020831180**

Product description **White respiratory protective half-face mask with valve, marked with FFP2 and with the CE marking printed on.**

Packaging description **Cardboard box containing 10 units, with a reference to EN 149 and the claim FFP2 printed on it.**

Is the product counterfeit? **Unknown**

Is this notification linked to a Business Gateway case?

Type of risk **Health risk / other**

Risk description (technical defect and consequences)

**The total filtration capacity of the mask is insufficient (measured mean values: as low as 75%). Consequently, the product does not fulfil the health and safety requirements; thus, even if combined with other recommended measures, it may not properly protect.**

Legal provisions (at EU level) and European standards against which the product was tested and did not comply

**The product does not comply with the Personal Protective Equipment (PPE) Regulation and with the European standard EN 149.**

(Overall) risk level **Serious**

Product category **Protective equipment**

Measures **Voluntary measures**

**Manufacturer**

**Destruction of the product  
31/08/2021**

Year

**2022**

Country of origin

**Lithuania**

Alert submitted by

**Latvia**