

Case number **A12/01195/22**

Product (what the product is) **Particle filter mask**

Name (on the product or the packaging) **FFP2 Filtrierende Atemschutzmaske**

Brand (on the product or the packaging) **• EN MED**

OECD Portal category (if known) **85000000 - Safety / Protection - DIY**

Is the product (also) sold online? **Unknown - No information**

Professional product? **No**

Product Recall

Type/number of model **ENM-712**

Batch number **20210309
20210307**

Bar code **8682431030478**

Product description **Protective respiratory filtration half-face mask marked with FFP2 and with the CE mark printed on.**

Packaging description **White-blue cardboard box containing 6 individually packed masks, with the claim FFP2 and the CE mark printed on.**

Is the product counterfeit? **Unknown**

Is this notification linked to a Business Gateway case?

Type of risk **Health risk / other**

Risk description (technical defect and consequences) **The particle/filter retention of the material is insufficient (measured values: as low as 56%). Consequently, the product does not fulfil the**

health and safety requirements; thus, even if combined with other recommended measures, it may not properly protect.

Legal provisions (at EU level) and European standards against which the product was tested and did not comply

The product does not comply with the Personal Protective Equipment Regulation nor with the European standard EN 149.

(Overall) risk level

Serious

Product category

Protective equipment

Measures

Voluntary measures

Retailer

**Recall of the product from end users
27/06/2022**

Compulsory measures

Other

None required

Voluntary measures

Importer

**No more goods
30/05/2022**

Year

2022

Country of origin

Turkey

Alert submitted by

Germany