
Case number **A12/01218/22**

Product (what the product is) **Particle filter mask**



Name (on the product or the packaging)

Brand (on the product or the packaging) **• Zen Mask**

OECD Portal category (if known) **85000000 - Safety / Protection - DIY**

Is the product (also) sold online? **Unknown - Not known**

Professional product? **No**

Product Recall

Type/number of model **FLS:001**

Batch number **MKZ-044-2021-9-6-8**

Bar code **8682966107034**

Product description **White particle filter mask with the claim FFP2, the CE marking and the words 'FLASH MED' printed on it.**

Packaging description **White cardboard box with blue lettering, containing 10 individually packaged masks.**

Is the product counterfeit? **Unknown**

Is this notification linked to a Business Gateway case?

Type of risk **Health risk / other**

Risk description (technical defect and consequences)

The particle/filter retention of the material is insufficient (measured values: as low as 55%). Consequently, the product does not fulfil the health and safety requirements; thus, even if combined with other recommended measures, it may not properly protect.

Legal provisions (at EU level) and European standards against which the product was tested and did not comply

The product does not comply with the Personal Protective Equipment Regulation nor with the European standard EN 149.

(Overall) risk level **Serious**

Product category **Protective equipment**

Measures **Voluntary measures**

Retailer

**Warning consumers of the risks
20/06/2022**

Year

2022

Country of origin

Turkey

Alert submitted by

Germany