

Case number **INFO/00089/22**

Product (what the product is) **Particle filter mask**



Name (on the product or the packaging) **PUREMEN KN95**

Brand (on the product or the packaging) **• PUREMEN**

OECD Portal category (if known) **85000000 - Safety / Protection - DIY**

Is the product (also) sold online? **No**

Professional product? **No**

Product Recall

Type/number of model **KN95 MASK-001**

Batch number **03/2023**

Bar code **8720246200324**

Product description **Protective respiratory filtration half-face mask. The mask is marked with both FFP2 and KN95 indications, and has the CE marking printed on the side.**

Packaging description **The product is sold in a white and blue plastic pouch, containing 2 items, and with user information printed on the back only in (presumably Chinese) characters. A label with the reference to**

standard EN 149, both claims FFP2 and KN95, and with the CE marking, is attached on the bottom rear side.

Is the product counterfeit? **Unknown**

Is this notification linked to a Business Gateway case?

Type of risk **Health risk / other**

Risk description (technical defect and consequences) **The product bears a CE marking but its filtering capacity has not been tested by a relevant competent European conformity assessment body. Consequently, it is not proven that the product fulfils the health and safety requirements; thus, even if combined with other recommended measures, it may not properly protect.**

Legal provisions (at EU level) and European standards against which the product was tested and did not comply **The product does not comply with the Personal Protective Equipment Regulation and with the European standard EN 149.**

(Overall) risk level **Unknown**

Product category **Protective equipment**

Measures **Voluntary measures**

Retailer

**Stop of sales at all locations
03/02/2022**

Year

2022

Country of origin

People's Republic of China

Alert submitted by

Germany