

Safety Gate RAPEX

Submission number

FI/00133/23

Case number

A12/00139/23

Product (what the product

is)

Particle filter mask





Name (on the product or the packaging)

Brand (on the product or the packaging)

• Lifa air

OECD Portal category (if known)

85000000 - Safety / Protection - DIY

Is the product (also) sold online?

Unknown - Not known.

Professional product?

No

Product Recall

Type/number of model

FFP2 NR LMp300x

Batch number

41/21

Bar code

6430077231412

Product description

Respiratory half mask, available in different colours, with CE marking.

Packaging description

Cardboard box, containing 9 items, with CE marking. The packaging

indicates manufacturing on 11.10.2021 and expiry on 11.10.2026.

Is the product counterfeit?

No

Is this notification linked to a Business Gateway case?

Type of risk

Health risk / other

Risk description (technical defect and consequences) The particle/filter retention of the material is insufficient (measured values as low as 46%). Consequently, the product does not fulfil the health and safety requirements; thus, even if combined with other recommended measures, it may not properly protect.

Legal provisions (at EU level) and European standards against which the product was tested and did not comply The product does not comply with the Personal Protective Equipment Regulation nor with the European standard EN 149.

(Overall) risk level

Serious

Product category

Protective equipment

Measures

Compulsory measures

Manufacturer

Withdrawal of the product from the market 16/11/2022

Compulsory measures

Manufacturer

Recall of the product from end users

16/11/2022

Compulsory measures

Manufacturer

Stop of sales 16/11/2022

Year

Country of origin

Alert submitted by

2023

Finland

Finland